

# CERTIFICATE OF CONFORMANCE

*Certificate #:* 2001230031

**Model:** QLA141

**Item #:** QSH00161

**Capacity:** 25 lb

**Temperature:** 71°F

**Humidity:** 38 %

**Excitation:** 4.59 (Vdc)

**FS Load:** 25 lb

## ***Tension***

No.	S/N	Rin ( $\Omega$ )	Rout ( $\Omega$ )	FS Output (mV/V)	Zero Balance (mV/V)
1		352	352	2.4999	0.0118

This certifies that the listed sensor(s) and/or instrument(s) has been tested using equipment traceable to NIST and conforms to the manufacturing and/or final product specifications per printed material or purchase order. Supporting documentation relative to traceability is on file and is available for examination upon request.

Tested by: **Vandy Say**

1/23/2020